



UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE
United States Patent and Trademark Office
Address: COMMISSIONER FOR PATENTS
P.O. Box 1450
Alexandria, Virginia 22313-1450
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/056,807	01/25/2002	Jon Occl	M190.134.101	9381
27581	7590	09/14/2007		
MEDTRONIC, INC. 710 MEDTRONIC PARKWAY NE MINNEAPOLIS, MN 55432-9924			EXAMINER VRETTAKOS, PETER J	
			ART UNIT 3739	PAPER NUMBER
			MAIL DATE 09/14/2007	DELIVERY MODE PAPER

Please find below and/or attached an Office communication concerning this application or proceeding.

The time period for reply, if any, is set in the attached communication.

Office Action Summary

Application No.

10/056,807

Applicant(s)

OCEL ET AL.

Examiner

Peter J. Vrettakos

Art Unit

3739

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --

Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) OR THIRTY (30) DAYS, WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

Status

- 1) ☒ Responsive to communication(s) filed on 10 August 2007.
- 2a) ☐ This action is **FINAL**. 2b) ☒ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

Disposition of Claims

- 4) ☒ Claim(s) 54 and 56-71 is/are pending in the application.
- 4a) Of the above claim(s) _____ is/are withdrawn from consideration.
- 5) ☐ Claim(s) _____ is/are allowed.
- 6) ☒ Claim(s) 54 and 56-71 is/are rejected.
- 7) ☐ Claim(s) _____ is/are objected to.
- 8) ☐ Claim(s) _____ are subject to restriction and/or election requirement.

Application Papers

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☐ The drawing(s) filed on _____ is/are: a) ☐ accepted or b) ☐ objected to by the Examiner.
- Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).
- Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

Priority under 35 U.S.C. § 119

- 12) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☐ All b) ☐ Some * c) ☐ None of:
- ☐ Certified copies of the priority documents have been received.
 - ☐ Certified copies of the priority documents have been received in Application No. _____.
 - ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).

* See the attached detailed Office action for a list of the certified copies not received.

Attachment(s)

- ☒ Notice of References Cited (PTO-892)
- ☐ Notice of Draftsperson's Patent Drawing Review (PTO-948)
- ☐ Information Disclosure Statement(s) (PTO/SB/08)
Paper No(s)/Mail Date _____.
- ☐ Interview Summary (PTO-413)
Paper No(s)/Mail Date. _____.
- ☐ Notice of Informal Patent Application
- ☐ Other: _____.

Art Unit: 3739

DETAILED ACTION

Claims 54 and 56-71 are pending.

54 and 71 are independent.

The action is non-final.

A new rejection is added below in light of the Applicant's arguments. Prior rejections are withdrawn.

Claim Rejections - 35 USC § 103

The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negated by the manner in which the invention was made.

Claims 54 and 56-71 are rejected under 35 U.S.C. 103(a) as being unpatentable over Wakikaido et al. (6,451,014) in view of Panescu et al. (6,245,065).

Wakikaido discloses a medical device for ablation (coagulation) with a rigidly coupled manually graspable handle (13), a metal (col. 6:66) malleable (as provided by spring member 30) elongated shaft (9) with a pivoting joint (20) with pins (26), non-conductive material (21, figure 2, col. 7:42-44), a slidable (see patented claim 1 – movable support 20 with electrode that can be bent and stretched in the direction of the longitudinal axis of the support shaft) rounded/uniform radius of curvature distal tip

Art Unit: 3739

portion (col. 6:57-63, "ball-like") with opening/hole (holding electrode 5), a remote actuator (slot 44), and a passage/internal lumen (to permit coaxial cable (10) or spring member (30)).

The patent discloses a slot/slide (42,44 in figure 6). Regarding claims 60-63, these are obvious design choices for actuators in light of the patented actuator (42,44). Further, the Applicant neglects to disclose criticality or unexpected results in choosing one actuator design over the other. (The Office views these actuators as obvious of each other and seamlessly substitutable.)

Wakikaido discloses no light connected to a power source.

Panescu discloses in an analogous ablation system, a light (217, col. 8:49-53) connected to a power source, as well as sensors and switches/activators (212 on the handle being an alternate to element 215; col. 8:31-39). (It would be obvious that lights 217 could be used in conjunction with handle switch 212).

Therefore, at the time of the invention it would have been obvious to one of ordinary skill in the art to modify Wakikaido in view of Panescu to include a light attached to the Wakikaido handle/medical device in order (motivation) to apprise the physician of the instantaneous on/off state of energy application.

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Peter J. Vrettakos whose telephone number is 571-272-4775. The examiner can normally be reached on M-F 9-6.


Art Unit: 3739

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Linda C. Dvorak can be reached on 571-272-4764. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.



Pete Vrettakos
September 6, 2007



ROY D. GIBSON
PRIMARY EXAMINER